

Sheet1

KTev 9954KB		Original B spec data											Fermilab Lab6		
S/No	A/No.	B.I. Date	Sens(360V)	Sens(180V)	CB	CR	D1 gain	500 A/L	DC	Leak	2000 A/L	DC	Grade	QE wrt Ref	Measurement
			μA/Lm	μA/Lm	μA/Lm	μA/Lm		V	nA	nA	V	nA			
8593	867920	10/01/92	69	68	8.6	3.3	7.4	1962	1.70	0.01	2173	5.00	PASS	53%	
8912	881439	10/30/92	81	80	8.6	3.3	6.9	1843	1.00	0.00	2137	4.00	PASS	58%	
9137	905101	10/20/93	71	71	12.3	1.7	6.7	2135	1.12	0.01	2498	3.00	PASS	34%	
9146	905223	10/20/93	74	74	11.8	1.9	5.1	2016	3.50	0.26	2298	16.00	PASS	37%	
9174	905170	10/21/93	74	73	12.6	1.7	8.5	2001	1.30	0.00	2342	9.20	PASS	46%	
9312	905472	11/09/93	73	73	9.2	1.3	5.2	2080	3.40	0.02	2378	17.00	PASS	53%	
9454	907023	11/24/93	75	74	12.4	1.8	7.1	2144	1.80	0.05	2480	7.00	PASS	37%	
9458	907036	11/24/93	75	73	12.0	1.9	7.3	2173	4.00	0.20	2500	7.00	PASS	42%	
9462	906118	11/25/93	80	79	10.7	2.4	5.0	2029	2.30	0.01	2308	5.50	PASS	48%	
9473	907022	11/25/93	72	72	11.9	2.4	6.5	2058	1.00	0.00	2325	2.50	PASS	45%	
8593	Retest	06/01/03	55		7.3	2.5		1981							